SRI VIVEKANANDA RESEARCH LABORATORIES

NABL Accredited Laboratory

(Regn. No. 290202201408 Dt. : 26/11/2013 Govt. of Karnataka)

22, 10th Cross, 1st Floor, Doddanna Indl. Estate, Bangalore - 560 091
Ph.: 080-22728531, Mobile : 9945371498 WhatsApp No.: 7760433054
E-mail : vivekanandaresearchlab@gmail.com ceo.vivekanandalab@gmail.com

Certificate No. TC-5046

TEST REPORT

Date:22.02.2024

Customer Name *M/s. Hospitime india*
284,Phase I,HSIIDC Industrial Area,Barhi
Haryana-131101

Test Report No.	7046/SVRL/CR/2023-24
ULR NO.	TC504624200000845F
Test Conducted	Chemical Analysis
Standard Reference	IS:9879-1998 (By Optical Emission Spectrometer)
Nature Of Sample	S.S Sample
Customer Reference	Dtd: 22.02.204
Date Of Reciept Of Sample	22.02.2024
Date Of Testing Of Sample	22.02.2024

Chemical Analysis Report

Sample Description	*Elements*	*Specification Limits*	*% By Weight*
S.S Sample	Carbon	0.080 Max.	0.042
Surgical Dressing Trays Or instrument trays	Manganese	2.00 Max.	1.017
	Silicon	1.00 Max.	0.522
	Nickel	8.0-12.0	8.643
	Chromium	18.0-20.0	18.782
	Sulphur	0.030 Max.	0.015
	Phosphorous	0.045 Max.	0.022

Material Confirmation : The Above Chemical Composition Confirm to **S.S-304.**

*******End of the Report*******

Tested By
Chemist

Authorised Signatory
Quality /Technical Manager
T. C. Shivaramanna

Note :
1. Test Results Only to Sample Submitted by the Party.
2. Test Report Shall not be reproduced Except in Full.
3. Samples will disposed after fifteen days from the date of issue of Test Report unless otherwise specified.
4. The report not be used as an evidence in the court of law and should not be used in any advertising media without permission in writting.