micro
Testing & Calibration Solutions
TESTING, CALIBRATION & CERTIFICATION SERVICES

Certificate No : T-7309

An ISO 9001:2015 CERTIFIED COMPANY

TEST REPORT

ULR NUMBER	TC730918000000070P			
Certificate No. :	MTC18L031T	ISSUE DATE :	29/11/2018	Page 1 of 1

TESTED For M/S	HOSPITIME INDIA BM-32, East , Shalimar Bagh , Delhi - 110088

RECIPT OF MATERIAL	29/11/2018	Customer Ref Number	NIL
SAMPLE DESC	Surgical Instrument	PROTOCOL	IS 8811 : 1988
SAMPLE SR. NO.	18L031	Temperature/ Humidity	22 °C/ 45-60 % RH

ELEMENTS & CONC. %

Element Name	Element	Observed Value %	As Per AISI 410
Iron	Fe	82.82	
Carbon	C	0.148	0.08 - 0.15
Silicon	Si	0.302	1.0 Max
Manganese	Mn	0.322	1.0 max
Sulphur	S	0.009	0.03 Max
Phosphorus	P	0.027	0.040 Max
Chromium	Cr	12.884	11.5 - 13.5
Nickel	Ni	0.267	1.0 Max

Sample Return : YES

Remarks : Confirmed As Per AISI 410 Chemically

* Marked Elements Are Not Covered Under NABL Scope

******End of Report******

Micro Testing & Calibration Solutions Auth. Sign.

Authorized By
Rahul Tanwar
(Technical Manager)

Format : F/QCD/06(Rev 00)

Note : 1) This certificate refers to the item tested/ calibrated and applicable parameter and result certificate is valid at the time of and under the stated conditions of testing/ Calibration. 2) Any complaint about the certificate should be communicated in writing within 7 days of issue of this certificate. 3) Laboratory Standards are traceable to National/International Standards. 4) This certificate is not to be reproduced wholly or in parts and cannot be used as evidence in a court of law and shall not be used in advertising media without written permission of Proprietor. 5) The liability of sample tested or instrument calibrated is limited to the invoiced amount. No liability will be accepted after the sample or instrument is taken back. 6) Endorsement of any product is neither inferred nor implied as the testing/calibration pertains only to the sample/instrument tested/calibrated. 7) While "Micro Testing & Calibration Solutions" has made their best endeavors to provide best and accurate and reliable information ,"Micro Testing & Calibration Solutions" is not responsible for any financial liability due to any act of omission or error made.

Address : #2280, Basement, Aravali Vihar, Sector-49, Faridabad-121001
Tel : 09646458212, 08427683447
Website : www.mtcsolutions.in
E-mail : mtcsolutions4u@gmail.com